

Instructions for use

miniRITE R

Oticon Opn S™
Oticon Opn Play™
Oticon Ruby



Made for
Apple iPhone | iPad | iPod

oticon
life-changing technology



WARNING: People younger than 18 should go to a doctor before using this.

People younger than 18 years old need specialized care, and using this without a medical evaluation may worsen impairment or disability. A hearing aid user who is younger than 18 should have a recent medical evaluation from a doctor, preferably an ear-nose-throat doctor (an ENT). Before using this, a doctor should determine that the use of a hearing aid is appropriate.

WARNING to Hearing Aid Dispensers:

You should advise a prospective hearing aid user to consult promptly with a doctor, preferably an ear specialist such as an ENT, before dispensing a hearing aid if you determine through inquiry, actual observation, or review of any other available information

concerning the prospective user, that the prospective user has any of the following conditions:

- Visible deformity of the ear, either congenital or traumatic
- Fluid, pus, or blood coming out of the ear within the previous 6 months
- Pain or discomfort in the ear
- History of excessive ear wax or suspicion that something is in the ear canal
- Dizziness, either recent or long-standing
- Sudden, quickly worsening, or fluctuating hearing loss within the previous 6 months
- Hearing loss or ringing (tinnitus) only in one ear or a noticeable difference in hearing between ears
- Audiometric air-bone gap equal to or greater than 15 dB at 500 Hz, 1000 Hz, and 2000 Hz

**WARNING to Hearing Aid Dispenser,
Outputs over 132 dB SPL:**

You should exercise special care in selecting and fitting a hearing aid with a maximum output that exceeds 132 dB SPL because it may impair the remaining hearing of the hearing aid user.

Caution: This is not hearing protection.

You should remove this device if you experience overly loud sounds, whether short or long-lasting. If you're in a loud place, you should use the right kind of hearing protection instead of wearing this device. In general, if you would use ear plugs in a loud place, you should remove this device and use ear plugs.

Caution: The sound output should not be uncomfortable or painful.

You should turn down the volume or remove the device if the sound output is uncomfortably loud or painful. If you consistently need to turn the volume down, you may need to further adjust your device.

Caution: You might need medical help if a piece gets stuck in your ear.

If any part of your hearing aid, like the eartip, gets stuck in your ear, and you can't easily remove it with your fingers, get medical help as soon as you can. You should not try to use tweezers or cotton swabs because they can push the part farther into your ear, injuring your eardrum or ear canal, possibly seriously.

Note: What you might expect when you start using a hearing aid

A hearing aid can benefit many people with hearing loss. However, you should know it will not restore normal hearing, and you may still have some difficulty hearing over noise. Further, a hearing aid will not prevent or improve a medical condition that causes hearing loss.

People who start using hearing aids sometimes need a few weeks to get used to them. Similarly, many people find that training or counseling can help them get more out of their devices.



If you have hearing loss in both ears, you might get more out of using hearing aids in both, especially in situations that make you tired from listening –for example, noisy environments.

Note: Tell FDA about Injuries, malfunctions, or other adverse events.

To report a problem involving your hearing aid, you should submit Information to FDA as soon as possible after the problem. FDA calls them “adverse events,” and they might include: skin irritation in your ear, injury from the device (like cuts or scratches, or burns from an overheated battery), pieces of the device getting stuck in your ear, suddenly worsening hearing loss from using the device, etc.

Instructions for reporting are available at <https://www.fda.gov/Safety/MedWatch>, or call 1-800-FDA-1088. You can also download a form to mail to FDA.

Note: Hearing loss in people younger than 18

- People younger than 18 should see a doctor first, preferably an ear-nose-throat doctor (an ENT), because they may have different needs than adults.
- The doctor will identify and treat medical conditions as appropriate.
- The doctor may refer the person to an audiologist for a separate test, a hearing aid evaluation.
- The hearing aid evaluation will help the audiologist select and fit the appropriate hearing aid.

A person who is younger than 18 years old with hearing loss should have a medical evaluation by a doctor, preferably an ENT, before buying a hearing aid. The purpose of a medical evaluation is to identify and treat medical conditions that may affect hearing but that a hearing aid won't treat on its own.

Following the medical evaluation and if appropriate, the doctor will provide a written statement that the hearing loss has been medically evaluated and the person is a candidate for a hearing aid. The doctor may refer the person to an audiologist for a hearing aid evaluation, which is different from the medical evaluation and is intended to identify the appropriate hearing aid.

The audiologist will conduct a hearing aid evaluation to assess the person's ability to hear with and without a hearing aid. This will enable the audiologist to select and fit a hearing aid for the person's individual needs. An audiologist can also provide evaluation and rehabilitation since, for people younger than 18, hearing loss may cause problems in language development and educational and social growth. An audiologist is qualified by training and experience to assist in the evaluation and rehabilitation of hearing loss in people younger than 18.

Model overview

This booklet is valid for the Oticon Opn S™, Oticon Opn Play™ and Oticon Ruby families in the following hearing aid models:

FW 8.0

- | | |
|---|---------------------------|
| ☐ Oticon Opn S 1 | GTIN: (01) 05714464013166 |
| ☐ Oticon Opn S 2 | GTIN: (01) 05714464013173 |
| ☐ Oticon Opn S 3 | GTIN: (01) 05714464013180 |

- | | |
|--|---------------------------|
| ☐ Oticon Opn Play 1 | GTIN: (01) 05707131374998 |
| ☐ Oticon Opn Play 2 | GTIN: (01) 05707131375001 |

FW 1.0

- | | |
|--|---------------------------|
| ☐ Oticon Ruby 1 | GTIN: (01) 05714464013210 |
| ☐ Oticon Ruby 2 | GTIN: (01) 05714464013227 |

The following speakers are available for the above models:

- ☐ Speaker 60
- ☐ Speaker 85
- ☐ Speaker 100 (Power Instrument)

- ☐ Power receiver mold speaker 100
- ☐ Power receiver mold speaker 105

Introduction to this booklet

This booklet gives you guidance on how to use and maintain your new hearing aid. Please read the booklet carefully, including the **Warnings** section. This will help you to get the most out of your new hearing aid.

Your hearing care professional has adjusted the hearing aid to meet your needs. If you have further questions, please contact your hearing care professional.

A hearing care professional (hearing aid professional, audiologist, ENT (ear, nose and throat) doctor, and hearing aid dispenser) is a person who is appropriately educated and has proven competency in professionally assessing hearing, selecting, fitting, and delivering hearing instruments and rehabilitation care to persons with hearing loss.

The education of the hearing care professional is in accordance with national or regional regulations.

| **About** | Start-up | Handling | Options | Tinnitus | Warnings | More info |

For your convenience, this booklet contains a navigation bar to help you easily navigate through the different sections.

IMPORTANT NOTICE

The hearing aid amplification is uniquely adjusted and optimized to your personal hearing capabilities during the hearing aid fitting performed by your hearing care professional.

Intended use

Intended use	The hearing aid is intended to amplify and transmit sound to the ear.
Indications for use	Bilateral or unilateral impaired hearing of sensorineural, conductive, or mixed type ranging from a slight (16dB HL*) to profound (95dB HL*) degree of hearing loss, with an individual frequency configuration.
Intended user	Person with hearing loss using a hearing aid and their caregivers. Hearing care professional responsible for adjusting the hearing aid.
Intended user group	Adults and children older than 36 months.
User environment	Indoor and outdoor.

Contraindications	Not suitable for infants below 36 months. Users of active implants must pay special attention when using the hearing aid. For more information read the Warnings section.
Clinical benefits	The hearing aid is designed to provide better speech understanding to help ease communication with the aim of improving quality of life.

* As specified by the American Speech-Language-Hearing Association, asha-org, using pure-tone average of 0.5, 1 and 2 kHz.

In short

Charging

Charge the hearing aid every night.

The LED on the hearing aid will turn RED/ORANGE when the hearing aid is charging and GREEN when fully charged.

It takes 3 hours to fully charge the hearing aid.

ON/OFF

The hearing aid turns ON when you remove it from the charger.

The hearing aid turns OFF when you place it in the charger.

IMPORTANT NOTICE

Ensure you always leave the charger connected to a power source when the hearing aid is seated in the charging port. Turning the charger OFF makes the hearing aid turn ON and start using battery power.

Table of contents

About

Your hearing aid, speaker and earpiece	16
Hearing aid	18

Start-up

Charging time	20
Battery performance	21
Turn hearing aid ON/OFF using charger	22
Turn hearing aid ON/OFF using push button	23
Battery low indication	24
Identify left and right hearing aid	25
Put on hearing aid	26

Handling

Cleaning	28
Replace standard earpieces	30
ProWax miniFit filter	32
Clean custom earpieces	34

Continues on next page

Table of contents

	Replace ProWax filter	35
	Storage of the hearing aid	36
	Flight mode	38
Options		
	Optional features and accessories	40
	Change volume	41
	Change program	42
	Mute	43
	Use hearing aid with iPhone and iPad	44
	Wireless accessories and other options	46
Tinnitus		
	Tinnitus SoundSupport™ (optional)	48
	Sound options and volume adjustments	50
	Limitation on use time	53
	Tinnitus SoundSupport warnings	56

Warnings

General warnings	58
------------------	----

More info


Troubleshooting guide	66
Water & dust resistant (IP68)	70
Conditions of use	71
Cell phone	74
Technical information	76
Warranty	86
Summary of relevant studies	87
Your individual hearing aid settings	88
Sound and LED indicators	90
Technical Data	94

Your hearing aid, speaker and earpiece




The hearing aid uses one of the following speakers:

**Standard
speakers**



**Power
Receiver
Mold
speakers**



The speakers uses one of the following earpieces:

Standard earpieces



☐ Open dome



☐ Bass dome, single vent



☐ Bass dome, double vent



☐ Power dome



☐ Grip Tip

Available in small and large, left and right, with or without vent.

Custom earpieces



☐ MicroMold*



☐ LiteTip*

Dome sizes



5 mm* 6 mm 8 mm 10 mm 12 mm**


* only as open dome for speaker 60

** not for open dome

*Can be made of different material
Please see details for replacing the dome
in section "Replace standard earpieces".

Hearing aid

What it is and does



Charging time

Fully charge your hearing aid before first time use.

Charge your hearing aid every night. That ensures you start your day with a fully charged hearing aid.

If your hearing aid is completely drained the normal charging time is:

3 hours	1 hour	0.5 hour
Fully charged	50% charged	25% charged

The charging process will automatically stop when the battery is fully charged. Leave the hearing aid in the charger when you do not use it. Remember to keep the charger connected to a power source.

Charging time may vary depending on the remaining capacity of the battery and between the left and right hearing aid.
For how to use your charger see the charger **Instructions** for use.

Battery performance

The daily battery performance varies depending on your individual use and hearing aid settings.

Streaming sound from e.g. TV, mobile phone or ConnectClip can influence daily battery performance.

If your hearing aid runs out of battery charge, place it in the charger for a re-charge. Additional usage time cannot be achieved by trying to restart the hearing aid.

For best charging condition, the room temperature should be between +10°C to +35°C (+50°F to +95°F).

If you experience that your hearing aid does not perform a full day, it might be time to have the battery replaced. Contact your hearing care professional.

Turn hearing aid ON/OFF using charger

Your hearing aid automatically turns **ON** when removed from the charger.

The hearing aid LED turns **GREEN** after 6 seconds, confirming that it is ready for use. You may hear a start-up jingle.


Your hearing aid automatically turns **OFF** when placed in the charger and starts charging. The LED turns **RED/ORANGE**.

IMPORTANT NOTICE

Ensure you always leave the charger connected to a power source when the hearing aid is seated in the charging port. Turning the charger OFF makes the hearing aid turn ON and start using battery power.

Turn hearing aid ON/OFF using push button

The hearing aid can be turned ON and OFF using the push button.



ON:

Press and hold the lower part of the push button for approx. 2 seconds.

Release the button and wait until the hearing aid LED turns GREEN.

The hearing aid is now turned ON. You may hear a short start-up jingle.

OFF:

Press and hold the lower part of the button for approx. 3 seconds until the hearing aid LED turns RED/ORANGE and the hearing aid plays 4 descending tones.

Release the push button and the hearing aid is turned OFF.

See overview "Sound and LED indications".

Battery low indication

When the battery is running low, you will hear three short beeps. This will leave you approximately two hours before the hearing aid stops working. The beeps will be repeated every 30 minutes. Just before the battery runs out you will hear four descending tones.

If streaming audio to your hearing aid (from e.g. TV or mobile phone) you have approximately one hour before the hearing aid stops working, when you hear the three short beeps. Stopping streaming will prolong the battery performance.



Three beeps

= The battery is running low



Four descending tones

= The battery has run out


Optional LED

Red/orange blinks, continuously repeated, to indicate battery low.

Identify left and right hearing aid

It is important to distinguish between the left and the right hearing aid, as they may be programmed differently.

You can find left/right color indicators on the hearing aid itself and on 60 and 85 speakers as shown. Indicators (either L or R) can also be found on 100 speakers and some earpieces.



Put on hearing aid


Step 1



Place the hearing aid behind your ear.


The speaker should always be used with an earpiece attached. Use only parts designed for your hearing aid.

Step 2



Hold the bend of the speaker wire between your thumb and index finger. The earpiece should point toward the opening of the ear canal.

Step 3




Gently push the earpiece into your ear canal until the speaker wire sits close to your head.

If the speaker has an ear grip, place it in the ear so it follows the contour of the ear.

Cleaning

The MultiTool contains a brush and wire loop for cleaning and removing earwax. If you need a new MultiTool, please contact your hearing care professional.



IMPORTANT NOTICE


The MultiTool has a built-in magnet. Keep the MultiTool at least 30 centimeters (1 foot) away from credit cards and other magnetically sensitive devices.

When handling your hearing aid, hold it over a soft surface to avoid damage if you drop it.

Clean the microphone openings

Use the brush of the MultiTool to carefully brush debris away from the openings. Carefully brush the surface around the opening.

Make sure that no parts of the MultiTool are squeezed into the microphone openings by force. This may damage the hearing aid.




IMPORTANT NOTICE

To clean the hearing aids, use a soft, dry cloth. The hearing aids must never be washed or immersed in water or other liquids.

Replace standard earpieces


The standard earpiece (dome and Grip Tip) should not be cleaned. If the earpiece is filled with wax, replace it with a new one. Grip Tip needs to be replaced at least once a month.

Step 1




Hold on to the speaker and pull off the earpiece.

Step 2



Insert the speaker exactly into the middle of the earpiece to obtain a secure attachment.

Step 3



Push firmly to ensure that the earpiece is fastened securely.


IMPORTANT NOTICE

If the earpiece is not on the speaker when removed from the ear, the earpiece may still be in the ear canal. Consult your hearing care professional for further instructions.

ProWax miniFit filter

The speaker has a white wax filter attached to the end where the earpiece is attached. The wax filter keeps wax and debris from damaging the speaker. Replace the filter when clogged, or if the hearing aid does not sound normal. Alternatively, contact your hearing care professional.

Remove the earpiece from the speaker before replacing the wax filter.




IMPORTANT NOTICE

Ensure you always use the same type of wax filter as was originally supplied with the hearing aids. If you are in doubt about the use or replacement of wax filters, contact your hearing care professional.


Replace ProWax miniFit filter

1. Tool




Remove the tool from the shell. The tool has two pins, one empty for removal and one with the new wax filter.

2. Remove



Push the empty pin into the wax filter in the speaker and pull it out.

3. Insert



Insert the new wax filter using the other pin, remove the tool, and throw it out.

Note:


If you use a mold or LiteTip, your hearing care professional must replace the wax filter in the speaker.

Clean custom earpieces

The earpiece should be cleaned regularly.

The earpiece has a white wax filter*. The filter keeps wax and debris from damaging the speaker.


Replace the filter when clogged, or if the hearing aid does not sound normal. Alternatively, contact your hearing care professional.



* Not all molds have a wax filter. Filter composition can vary by mold. For more information contact your hearing care professional.


Replace ProWax filter

1. Tool




Remove the tool from the shell. The tool has two pins, one empty for removal and one with the new wax filter.

2. Remove



Push the empty pin into the wax filter in the earpiece and pull it out.

3. Insert



Insert the new wax filter using the other pin, remove the tool, and throw it out.

Storage of the hearing aid

The charger is the best storage for the hearing aid. Anytime your hearing aid is not in use, place it in the charger for charging. This will ensure your hearing aid is always charged.

To ensure the longest life of the rechargeable battery in the hearing aid do not expose to excessive heat. For example, do not leave the hearing aid in the sun in a window or in a car, even if the hearing aid is in the charger.

IMPORTANT NOTICE

Always leave the charger connected to a power source when the hearing aid is seated in the charging port. Switching off the charger will make the hearing aid turn on and start using power.

Long term storage (weeks, months)


Before storing the hearing aid for a longer period (more than 14 days) fully charge the hearing aid. Then turn the hearing aid off. This way the battery can be charged again.

It is necessary to fully charge the hearing aid every 6 months to protect the rechargeable battery. If a stored hearing aid is not charged within a 6 month time frame, the rechargeable battery will need to be replaced.

To see how to place the hearing aid in the charger see the Charger Instructions for use.

Flight mode

To activate and deactivate flight mode press and hold the push button (lower part) for 7 seconds.



Beep: The hearing aid plays 4 descending beeps followed by 4 beeps. This indicates that flight mode is activated or deactivated.

LED flight mode activated: A long red/orange blink followed by green, red/orange, red/orange. The LED will blink green, red/orange, red/orange a few times. This confirms that flight mode is activated.

LED flight mode deactivated: A long red blink followed by two long green blinks. This confirms that flight mode is deactivated.

See overview "Sound and LED indications".

Pressing the lower part of the push button on one hearing aid will activate flight mode on both hearing aids.

When flight mode is activated, Bluetooth® is turned off. The hearing aid will still be working.

Optional features and accessories


The features and accessories described in the following pages are optional. Please contact your hearing care professional to find out how your hearing aid is programmed.

If you experience difficult listening situations, a special program may be helpful. These are programmed by your hearing care professional.

Write down hearing situations in which you may need help.

Change volume


The push button allows you to adjust the volume.
You will hear a beep when you turn the volume up or down.



Change program

Your hearing aid can have up to four different programs. These are programmed by your hearing care professional. You will hear one to four tones when you change program depending on the program.

See the Sound and LED indicator section.



Press the push button to switch between programs.

The program cycle switches one program forward, for example program 1 to 2 or program 4 to 1.

Mute

Use the mute function if you need to silence the hearing aid. Your hearing aid can be muted by using one of the following optional devices:

Oticon ON app

ConnectClip

Remote Control 3.0

How to unmute your hearing aid

The hearing aid can be unmuted using one of the optional devices or by applying a short press to the upper or lower part of the button on the hearing aid.

IMPORTANT NOTICE

Do not use the mute function as an off switch, as the hearing aids are still using battery power in this mode.

Use hearing aid with iPhone and iPad

Your hearing aid is Made for iPhone® and allows for direct communication and control with iPhone, iPad® or iPod touch®. For assistance in using these products with your hearing aid, please contact your hearing care professional.



To see how to pair your hearing aid with your iPhone please visit www.oticon.com/support

For information on compatibility, please visit www.oticon.com/support/compatibility

Use of the Made for Apple badge means that an accessory has been designed to connect specifically to the Apple product(s) identified in the badge, and has been certified by the developer to meet Apple performance standards. Apple is not responsible for the operation of this device or its compliance with safety and regulatory standards. Please note that use of this Accessory with iPod, iPhone, or iPad may affect wireless performance.

Wireless accessories and other options

As an enhancement to your wireless hearing aid, a range of accessories is available. These can enable you to hear and communicate better in many everyday situations.

ConnectClip

When ConnectClip is paired with your mobile phone, you can use the hearing aid as a hands-free headset. ConnectClip can also be used as a remote microphone

TV Adapter 3.0

TV Adapter streams sound directly from a TV or electronic audio device to your hearing aid.

Remote Control 3.0

Offers the ability to change program, adjust volume, or mute your hearing aid.

Oticon ON app

An application that lets you control your hearing aid from your mobile phone or tablet. For iPhone, iPad, iPod touch, and Android devices. Ensure that you only download and

install Oticon ON app from the official app stores.

Telecoil program

Telecoil helps you hear better when using a telephone with a built-in loop or when you are in buildings with teleloop systems such as theaters, churches, or lecture rooms. This symbol or a similar sign is shown wherever a teleloop has been installed.



CROS - optional for Oticon Opn S 1, Opn S 2, Oticon Opn Play 1, Oticon Ruby 1 and Ruby 2

Solution for people with an unaidable hearing loss in one ear. Oticon CROS on the poorer ear transmits sound to the hearing aid on the better ear.

For more information visit www.oticon.com or contact your hearing care professional.

☐ **Tinnitus SoundSupport™** (optional)

Intended use of Tinnitus SoundSupport

Tinnitus SoundSupport is a tool intended to generate sounds to provide temporary relief for patients suffering from tinnitus as part of a tinnitus management program.

The target population is the adult population (over 18 years old).

Tinnitus SoundSupport is targeted to licensed hearing care professionals (audiologists, hearing aid specialists, or otolaryngologists) who are familiar with the evaluation and treatment of tinnitus and hearing loss. Fitting of Tinnitus SoundSupport must be done by a hearing care professional participating in a tinnitus management program.

Guidelines for Tinnitus SoundSupport Users

These instructions contain information about Tinnitus SoundSupport, which may have been enabled in your hearing aids by your hearing care professional.

Tinnitus SoundSupport is a tinnitus management device intended to generate sound of sufficient intensity and bandwidth to help manage tinnitus.

Your hearing care professional will also be able to offer the appropriate follow-up care. It is important to follow his/her advice and directions regarding such care.

Prescription use only

Good health practice requires that a person reporting tinnitus have a medical evaluation by a licensed ear physician before using a sound generator. The purpose of such an evaluation is to ensure that any medically treatable condition that may cause tinnitus is identified and treated prior to using a sound generator.

Sound options and volume adjustments

Tinnitus SoundSupport is programmed by your hearing care professional to match your hearing loss and preferences for tinnitus relief. It offers a number of different sound options. Together with your hearing care professional, you can select your preferred sound(s).

Tinnitus SoundSupport programs

Together with your hearing care professional you decide for which programs you may want to have Tinnitus SoundSupport activated. The sound generator can be activated in up to four different programs.

Mute

If you are in a program for which Tinnitus SoundSupport is activated, the mute functionality will mute only the environmental sounds, and not the sound from Tinnitus SoundSupport. See chapter: "Mute".

Volume adjustments with Tinnitus SoundSupport

When you select a hearing aid program for which Tinnitus SoundSupport is activated, your hearing care professional can only set the push button on your hearing aid to work as a volume control for the tinnitus relief sound.

Your hearing care professional will set the volume control for the sound generator in one of two ways:

- A) Change volume in each ear separately, or
- B) Change volume in both ears simultaneously.

miniRITE R

A) How to change Tinnitus SoundSupport volume in each ear separately

To **increase** volume (on one hearing aid only), use a short press on the upper part of the push button repeatedly until desired level is reached.
To **decrease** volume (on one hearing aid only), use a short press on the lower part of the push button repeatedly until desired level is reached.

B) How to change Tinnitus SoundSupport volume in both ears simultaneously

You can use one hearing aid to increase/decrease the sound in both hearing aids. When changing the volume in one hearing aid, the volume on the other hearing aid will follow.

To **increase** volume, use a short press on the upper part of the push button repeatedly.

To **decrease** volume, use a short press on the lower part of the push button repeatedly.

To be filled out by your hearing care professional.

Limitation on use time

Daily use

The volume levels of Tinnitus SoundSupport can be set to a level which could lead to permanent hearing damage when used for a prolonged period of time. Your hearing care professional will advise you of the maximum amount of time per day you should use Tinnitus SoundSupport. It should never be used at uncomfortable levels.

See table "Tinnitus SoundSupport: Limitation on use - Your individual hearing aid settings" at the end of this booklet to learn how many hours per day you can safely use the relief sound in your hearing aids.

Important information for hearing care professionals about Tinnitus SoundSupport

Device description

Tinnitus SoundSupport is a module function that can be enabled in the hearing aids by the hearing care professional.

Maximum wearing time

The wearing time of Tinnitus SoundSupport will decrease as you increase the level above 80 dB(A) SPL. The fitting software will automatically display a warning when the hearing aid exceeds 80 dB(A) SPL. See “Max wearing time indicator” next to the tinnitus fitting graph in the fitting software.

The volume control is deactivated

By default the volume control for the sound generator is deactivated in the hearing aid. Risk of noise exposure increases when the volume control is activated.

If the volume control is activated

A warning may be displayed if you activate the tinnitus volume control in the “Buttons & Indicators” screen. This occurs if the relief sound can be listened to at levels that may cause hearing damage. The “Max wearing time” table in the fitting software displays the number of hours the patient can safely use Tinnitus SoundSupport.

- Note the max wearing time for each program for which Tinnitus SoundSupport is activated.
- Write those values in the table: “Tinnitus SoundSupport: Limitation on use”, in the back of this booklet.
- Instruct your patient accordingly.

Tinnitus SoundSupport warnings

If your hearing care professional has activated the sound generator Tinnitus SoundSupport, please pay attention to the following warnings.

There are some potential concerns associated with the use of any sound generated by a tinnitus management device. Among them are the potential worsening of tinnitus, and/or a possible change in hearing thresholds.

Should you experience or notice a change in hearing or tinnitus, or any dizziness, nausea, headaches, heart palpitations, or possible skin irritation at the point of contact with the device, you should immediately discontinue use of the device and consult a medical, audiology, or other hearing care professional.

As with any device, misuse of the sound generator feature may cause potentially harmful effects. Care should be taken to prevent unauthorized use and to keep the device out of reach of children and pets.

Maximum wearing time

Always follow the maximum wearing time per day of the Tinnitus SoundSupport advised by your hearing care professional. Prolonged use may lead to worsening of your tinnitus or of your hearing loss.

General warnings

For your personal safety and to ensure correct usage, you should familiarize yourself fully with the following general warnings before using your hearing aids.

Consult your hearing care professional if you experience unexpected operations or serious incidents with your hearing aids during use or because of its use. Your hearing care professional will support you with issue handling and, if relevant, reporting to the manufacturer and/or the national authorities.

Note that hearing aids do not restore normal hearing and do not prevent or improve a hearing impairment resulting from organic conditions. Hearing aids are only a part of hearing habilitation and may need to be supplemented by auditory training and instruction in lipreading. Furthermore, note that in most cases, infrequent use of hearing

aids does not permit a user to attain full benefit from it.

This hearing aid is supported by a non-removable rechargeable lithium-ion battery cell. Please ensure to charge the hearing aid and familiarize yourself with the safety and handling information related to rechargeable hearing aids.

Do not try to get access to the battery inserted in the hearing instrument. The battery must only be replaced by your hearing care professional. Only charge the hearing aids with a designated charger. Other chargers risk destroying the hearing aids and batteries.

If a battery or hearing aid is swallowed, see a doctor immediately.

Usage of hearing aids

Hearing aids should be used only as directed and adjusted by your hearing care professional. Misuse can result in sudden and permanent hearing loss.

Never allow others to wear your hearing aid, as incorrect usage could cause permanent damage to their hearing.

Choking hazards & risk of swallowing small parts

Hearing aids and their parts should be kept out of reach of children and anyone who might swallow these items or otherwise cause injury to themselves.

If a hearing aid or small part is swallowed, see a doctor immediately.

Explosives

The hearing aid is safe to use under normal usage conditions. The hearing aid has not been tested for compliance with international standards concerning explosive environments.

Therefore, do not use the hearing aid in environments with danger of explosions e.g. mines, oxygen rich environments or areas where flammable anaesthetics are handled.

General warnings

Fatality hazards and risk of swallowing Lithium-ion batteries or placing them in the ear or nose

Never swallow Lithium-ion batteries. Do not place them in the ear or the nose as this may lead to serious injury or death in as little as two hours. This can be due to chemical burn, which can cause permanent damage to the nose or ear or potentially lead to perforation of the inner organs. If a Lithium-ion battery is swallowed or placed in the ear or nose, seek emergency medical treatment immediately.

Keep the batteries in the original packaging until use. Dispose of used batteries immediately.

If a battery or hearing aid is swallowed, see a doctor immediately and contact the National Poison Center at 1-800-222-1222 or National Battery Ingestion Hotline at 1-800-498-8666.

Rechargeable battery

Do not attempt to open the hearing aid, as it may damage the battery.

Never attempt to replace the battery. If battery replacement is needed, return your device to the supplier. The service guarantee is void if there are signs of tampering.

In case of battery leakage, do not wear or carry your hearing aids, as it may cause skin irritation due to acids leaking from the battery.

If your skin has been in contact with the leaked battery acids, use a wet cloth to wipe it off and ensure no acid is left on your skin. If you experience skin irritation, consult your doctor.

For further handling instructions of your hearing aids consult your hearing care professional.

The safety of recharging batteries using a USB connector is determined by the external signal source. When connected to external equipment plugged into a power socket, this equipment must comply with IEC 62368-1 or equivalent safety standards.

Dysfunction

Be aware of the possibility that your hearing aids may stop working without notice. Keep this in mind when you depend on warning sounds (e.g. when you are in traffic). The hearing aids may stop functioning for instance if the batteries have expired or if the tubing is blocked by moisture or earwax.

Active implants

The hearing aid has been thoroughly tested and characterized for human health according to international standards for human exposure (Specific Absorption Ratio - SAR), induced

electromagnetic power and voltages into the human body.

The exposure values are well below internationally accepted safety limits for SAR, induced electromagnetic power and voltages into the human body defined in the standards for human health and coexistence with active medical implants, such as pacemakers and heart defibrillators.

If you have an active brain implant, contact the manufacturer of your implantable device for information about the risk of disturbance.

The AutoPhone magnet or MultiTool (which has a built-in magnet) should be kept more than 30 centimeters (1 foot) away from the implant, e.g. do not carry it in your breast pocket.

Follow the guidelines recommended by the manufacturers of implantable

Continues on next page

General warnings

defibrillators and pacemakers regarding their use with magnets.

Cochlear implants

If you are using a cochlear implant (CI) on one ear and a hearing aid on the other ear, make sure to always keep your CI at least at a 1 centimeter (0.4 inches) distance from your hearing aid. The magnetic field from the CI sound processors, coils, and magnets may permanently damage the speaker unit in your hearing aid. Never place the devices close together on a table e.g. when cleaning or changing batteries. Do not carry the CI and the hearing aid together in the same box.

X-ray/CT/MR/PET scanning, electrotherapy and surgery

Remove your hearing aid before X-ray examinations and CT/MR/PET scans,

electrotherapy, surgery, etc. as your hearing aid may be damaged when exposed to strong electromagnetic fields.

Heat and chemicals

The hearing aid must never be exposed to extreme heat, e.g. left inside a parked car in the sun.

The hearing aid must not be dried in microwave ovens or other ovens.

The chemicals in cosmetics, hairspray, perfume, aftershave lotion, sunscreen lotion, and insect repellent can damage the hearing aid. Always remove your hearing aid before applying such products and allow time to dry before use.

Detached earpiece in ear canal

If the earpiece is not on the speaker when removed from the ear, the earpiece may still be in the ear canal. For further instructions, consult your hearing care professional.

Power hearing aid

Special care should be exercised in selecting, fitting, and using hearing aids where the maximum sound pressure capability exceeds 132 dB SPL (IEC 60318-4/IEC 711) as there may be a risk of impairing the remaining hearing of the hearing aid user.

For information on whether your instrument is a power hearing aid, see the model overview section in this booklet.

Possible side effects

Hearing aids, molds or domes may cause an accelerated accumulation of earwax.

The non-allergenic materials used in hearing aids may in rare cases cause a skin irritation or other side effects.

Please seek consultation with a physician if these conditions occur.

Use on aircraft

Your hearing aids have Bluetooth wireless technology. On board an aircraft, the hearing aid must be put into flight mode to deactivated Bluetooth, unless Bluetooth is permitted by the flight personnel.

Connection to external equipment

The safety of the use of hearing aids when connected to external equipment (with an auxiliary input cable and/or USB cable and/or directly), is determined by the external signal source. When connected to external equipment plugged into a wall outlet, this equipment must comply with IEC 62368-1 or equivalent safety standards.

Use of third-party accessories

Only use accessories, transducers or cables supplied by the manufacturer. Non-original accessories may result in reduced electromagnetic compatibility (EMC) of your hearing aids.

Modification of hearing aids is not allowed

Changes or modifications not expressly approved by the manufacturer will void the warranty of the equipment.

Interference

The hearing aid has been thoroughly tested for interference according to the most stringent international standards.

Electromagnetic interference may occur in the vicinity of equipment with the symbol to the left. Portable and mobile RF (radio frequency) communications equipment can affect the performance of the hearing aid. If your hearing aid is affected by electromagnetic interference, move away from the source to reduce the interference

Troubleshooting guide

Troubleshooting for Charger, see Charger Instructions for use

Symptom	Possible causes
No sound	Hearing aid is out of power
	Dead battery
	Clogged earpieces (dome, Grip Tip, or mold)
Intermittent or reduced sound	Clogged sound outlet
	Moisture
	Hearing aid is out of power
Squealing noise	Hearing aid earpiece inserted incorrectly
	Earwax accumulated in ear canal
Beeping	If your hearing aid plays 8 beeps, 4 times consecutively, your hearing aid needs a microphone service check

If none of the above solutions work, consult your hearing care professional.

Solutions

Charge the hearing aid

Contact your hearing care professional

Clean mold

Replace wax filter, dome, or Grip Tip

Clean mold or replace wax filter, dome, or Grip Tip

Gently wipe the hearing aid and let it dry

Charge the hearing aid

Re-insert the earpiece

Have ear canal examined by your doctor

Contact your hearing care professional

Continues on next page

Troubleshooting guide

Troubleshooting for Charger, see Charger Instructions for use

Symptom	Possible causes
Hearing aid LED remains off when the hearing aid is placed in the charger	The charger is not switched on
	The hearing aid or charger is either too warm or too cold
	Charging is incomplete. The room temperature exceeds +95°F, which prolongs the charging time. The charger has stopped charging to protect the battery
	The hearing aid is not correctly inserted into the charger
The hearing aid LED blinks red/orange when the hearing aid is placed in the charger	System error
Pairing issue with Apple device	Bluetooth connection failed
	Only one hearing aid paired

Solutions

Verify that the charger power supply is connected correctly

Move the charger and hearing aid to a location with a temperature between +41°F to +104°F

Reinsert the hearing aid into the charger. This will complete the charging within approximately 15 minutes

Check the charger slots for foreign objects

Contact your hearing care professional

- 1) Unpair your hearing aid
- 2) Turn Bluetooth off and on again.
- 3) Turn off and turn back on the hearing aid.
- 4) Re-pair hearing aid (please visit www.oticon.com/support).

If none of the above solutions work, consult your hearing care professional.

Water & dust resistant (IP68)

Your hearing aid is dust tight and protected against ingress of water, which means it is designed to be worn in all daily life situations. Therefore, you do not have to worry about sweat or getting wet in the rain.

Before charging the hearing aid make sure to wipe off any moisture.

Should your hearing aid come into contact with water and stop working, gently wipe off any water and let the hearing aid dry.

IMPORTANT NOTICE

Do not wear your hearing aid while showering or participating in water activities. Do not immerse your hearing aid in water or other liquids.

Conditions of use

Operating conditions	Temperature: +5°C to +40°C (41°F to 104°F) Humidity: 5% - 93% relative humidity, noncondensing Atmospheric pressure: 700 hPa to 1060 hPa
Charging conditions	Temperature: +5°C to +40°C (41°F to 104°F) Humidity: 5% - 93% relative humidity, noncondensing Atmospheric pressure: 700 hPa to 1060 hPa

Continues on next page

Transportation and storage conditions	Temperature and humidity should not exceed the following limits for extended periods during transportation and storage: Transportation: Temperature: -20°C to + 60°C (-4°F to 140°F) (rechargeable hearing instruments) Humidity: 5% - 93% relative humidity, noncondensing Atmospheric pressure: 700 hPa to 1060 hPa Storage: Temperature: -20°C to + 30°C (-4°F to 86°F) Humidity: 5% - 93% relative humidity, noncondensing Atmospheric pressure: 700 hPa to 1060 hPa
--	--

Information about Charger: Condition of use, see Charger Instruction for use

Cell phone

Some hearing aid users have reported a buzzing sound in their hearing aid when using cell phones, indicating that the cell phone and hearing aid may not be compatible.

The ANSI C63.19 standard determines the prediction of compatibility between a specific hearing aid and a cell phone, thus hearing aid compliance is tested according to this standard. However, demonstrating compliance according to this standard cannot guarantee that all users will be satisfied.

Whereas all hearing aids have acoustic coupling, only the larger hearing aids have the physical space for telecoil (inductive) coupling.

The hearing aid is compliant with ANSI C63.19 in both microphone and telecoil mode.

IMPORTANT NOTICE

The performance of individual hearing aids may vary with individual cell phones. Therefore, ensure you try this hearing aid with your cell phone or, if you are purchasing a new phone, be sure to try it with your hearing aid prior to purchase. For additional guidance, please ask your cell phone provider for the booklet entitled "Hearing Aid Compatibility with Digital Wireless Cell Phones."

Technical information

The hearing aids contain the following two radio technologies:

The hearing aids contain a radio transceiver using short range magnetic induction technology operating at 3.84 MHz. The magnetic field strength of the transmitter is very weak and always below 15 nW [typically below $-40 \text{ dB}\mu\text{A/m}$ at a distance of 10 meters ($-12.20 \text{ dB}\mu\text{A/ft}$) at a distance of 33 feet].

The hearing aids also contain a radio transceiver using Bluetooth Low Energy technology and a proprietary, short-range radio technology both operating at ISM band 2.4 GHz.

The radio transmitter is weak and always below 3 mW equal to 4.8 dBm in total radiated power.

Only use your hearing aids in areas where wireless transmission is permitted.

The hearing aids comply with international standards concerning radio transmitters, electromagnetic compatibility, and human exposure.

Due to the limited space available on the hearing aid, relevant approval markings can be found in this booklet.

Additional information can be found in the "Technical Data sheets" on www.oticon.com

This device contains a radio module with the following certification ID numbers: FCC ID: 2ACAHSBMTRC

Radiofrequency radiation exposure information

This device complies with FCC RF exposure limits set forth for an uncontrolled environment and has been tested for portable use.

The device must not be co-located or used in conjunction with any other antenna or transmitter.

Use of other accessories not verified by the manufacturer may not ensure compliance with FCC RF exposure guidelines.

Note: This device has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This device generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this device does cause harmful interference to radio or television reception, which can be determined by turning the equipment

off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the manufacturer or an experienced radio/TV technician for help.

Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.



SBO Hearing A/S
Kongebakken 9
DK-2765 Smørum
Denmark

Should your hearing aid require service or replacement, contact your hearing care professional for assistance. Many repair needs can be handled on-site at your local hearing care professional's office, and they will arrange for service with the manufacturer if required. You can also contact us at: 580 Howard Ave., Somerset, NJ 08873.

CE 0123



Waste from electronic
equipment must be
handled according to
local regulations.

 **Bluetooth®**



IP68

Description of symbols used in this booklet



Warnings

Text marked with a warning symbol must be read before using the device.



Manufacturer

The device is produced by the manufacturer whose name and address are stated next to the symbol. Indicates the medical device manufacturer, as defined in EU Regulation 2017/745.



CE mark

The device complies with all required EU regulations and directives. The four digit number indicates the identification of the notified body.



Electronic waste (WEEE)

Recycle hearing aids, accessories or batteries according to local regulations. Hearing aid users can also return electronic waste to their hearing care professional for disposal. Electronic equipment covered by Directive 2012/19/EU on waste and electrical equipment (WEEE).



Regulatory Compliance Mark (RCM)

The device complies with electrical safety, EMC and radio spectrum requirements for devices supplied to the Australian or New Zealand markets.



IP code

Class of protection against harmful ingress of water and particulate matter according to EN 60529. IP6X indicates total dust protection. IPX8 indicates the protection against the effects of continuous immersion in water.

Description of symbols used in this booklet



Bluetooth

Bluetooth logo

Registered trademark of Bluetooth SIG, Inc. where any use of such requires a license.



Made for Apple badges

The device is compatible with iPhone, iPad and iPod touch.



Hearing loop

This logo incorporates the universal symbol for hearing assistance. The "T" signifies that a hearing loop is installed.



Radio Frequency (RF) transmitter

Your hearing aid contains an RF transmitter.

GTIN

Global Trade Item Number

A globally unique 14-digit number used to identify medical device products including medical device software. GTIN in this booklet is related to medical device firmware (FW). GTIN on regulatory packaging label is related to medical device hardware.

Description of additional symbols used on labels



Keep dry

Indicates a medical device that needs to be protected from moisture.



Catalog number

Indicates the manufacturer's catalog number so that the medical device can be identified.



Serial number

Indicates the manufacturer's serial number so that a specific medical device can be identified.



Medical Device

The device is a medical device.



Li-ion

Battery recycling symbol

Li-Ion Battery recycling symbol



Temperature limit

Indicates the temperature limits to which the medical device can be safely exposed

Description of additional symbols used on labels



Humidity limitation

Indicates the range of humidity to which the medical device can be safely exposed.



Unique device identifier

Indicates a carrier that contains unique device identifier information.

International warranty

Your hearing aid is covered by an international limited warranty issued by the manufacturer for a period of 12 months from the date of delivery to you. This limited warranty covers manufacturing and material defects in the hearing aid itself, but not in accessories such as batteries, tubing, speakers, earpieces, filters, etc. Problems arising from improper/incorrect handling or care, excessive use, accidents, repairs made by an unauthorized party, exposure to corrosive conditions, physical changes in your ear, damage due to foreign objects entering the device, or incorrect adjustments are NOT covered by the limited warranty and may void it. Under this warranty, the manufactur-

er will choose, at its sole discretion, whether to repair the hearing aid, or replace it with an equivalent model. The above warranty does not affect any legal rights that you might have under applicable national legislation governing the sale of consumer goods. Your hearing care professional may have issued a warranty that goes beyond the clauses of this limited warranty. Consult him/her for further information.

If you need service

Take your hearing aid to your hearing care professional, who may be able to solve minor problems and make adjustments immediately.

Your hearing care professional can assist you in obtaining warranty service from the manufacturer. Your hearing care professional may charge a fee for their services.

Warranty

Certificate

Name of owner: _____

Hearing care professional: _____

Hearing care professional's address: _____

Hearing care professional's phone: _____

Purchase date: _____

Warranty period: _____ Month: _____

Model left: _____ Serial no.: _____

Model right: _____ Serial no.: _____

Summary of relevant studies

Clinical evaluations conducted by or for the manufacturer provide evidence to support the intended use and clinical benefits outlined in the IFU and demonstrate regulatory conformity. Clinical data is collected, assessed, and analyzed to support the performance of the hearing aids by validating that they provide sufficient audibility and hearing loss compensation based on best-practice prescriptive fitting rationales. The clinical data also demonstrate improved speech understanding and success with hearing aids using validated questionnaires and surveys.

Non-clinical data supporting the overall performance of the hearing aids includes software verification, electroacoustic verification, electrical and mechanical safety evaluation, electromagnetic compatibility (EMC) evaluation, and documentation of radio properties and performance. Additional information can be found in section Technical Information.

Your individual hearing aid settings

To be filled out by your hearing care professional.

Tinnitus SoundSupport: Limitation on use			
☐	No limitation on use		
	Program	Start-up volume (Tinnitus)	Max volume (Tinnitus)
☐	1	Max _____ hours per day	Max _____ hours per day
☐	2	Max _____ hours per day	Max _____ hours per day
☐	3	Max _____ hours per day	Max _____ hours per day
☐	4	Max _____ hours per day	Max _____ hours per day

Settings overview for your hearing aid				
Left			Right	
☐ Yes	☐ No	Volume control	☐ Yes	☐ No
☐ Yes	☐ No	Program shift	☐ Yes	☐ No
☐ Yes	☐ No	Tinnitus SoundSupport	☐ Yes	☐ No
Volume control indicators				
☐ On	☐ Off	Beeps at min/max volume	☐ On	☐ Off
☐ On	☐ Off	Beeps when changing volume	☐ On	☐ Off
☐ On	☐ Off	Beeps at start-up volume	☐ On	☐ Off
Battery indicators				
☐ On	☐ Off	Low battery warning	☐ On	☐ Off

Sound and LED indicators

Different sounds and LED lights indicate the hearing aid status. The different indicators are listed on the following pages. For charger LED indicators, see Charger Instruction for use.

Your hearing care professional can set sound and LED indicators to match your preferences.

Program	☐ Sound	☐ LED ¹⁾	When to use
1	1 tone		
2	2 tones		
3	3 tones		
4	4 tones		

 Green flash

1) LED continuous or repeated three times with small pauses

ON/OFF	Sound	LED	LED comments
On	☐ Jingle	☐  	Shown one time
Off	☐ 4 descending tones	☐ 	
Volume	Sound	LED	
Start-up volume	☐ 2 beeps	☐ 	
Minimum/maximum volume	☐ 3 beeps	☐ 	
Volume up/down	☐ 1 beep	☐ 	
Mute activated on ON app, ConnectClip or Remote Control 3.0		☐  	Continuous or repeated three times

 Long green flash
  Green flash
  Red/orange flash
  Long red/orange flash

Continues on next page

Accessories	 Sound	 LED	LED comments
TV Adapter	2 different tones	 	Continuous or repeated three times
ConnectClip remote microphone	2 different tones	 	
Flight mode	Sound	 LED	
Flight mode active	4 descending tones + short jingle	   	
Flight mode inactive	4 descending tones + short jingle	   ¹⁾	

 Long green flash
  Green flash
  Red/orange flash

1) Only available when three-time repetition is selected

Warnings	Sound	LED	LED comments
Low battery	 3 alternate tones	      	Continuously flashing
Battery shut down	4 descending tones		
Microphone service check needed	8 beeps repeated 4 times	    	Repeated four times
No light in the hearing aid LED when placed in the charger		Turned off	See Troubleshooting, solutions
The hearing aid LED flashes red/orange when the hearing aid is placed in the charger		  	Continuously flashing. See Troubleshooting, solutions

 Red/orange flash  Long red/orange flash

Technical Data

miniRITE R

60

Measured according to American National Standard ANSI S3.22-2014

and ANSI S3.55-2014/Part 5

Supply voltage: Lithium ion

0 dB SPL ref. 20 µPa		Oticon Opn S 1 Opn Play 1	Oticon Opn S 2 Ruby 1	Oticon Opn S 3 Opn Play 2 Ruby 2
OSPL90	Peak	105 dB SPL	105 dB SPL	105 dB SPL
	HF Average	102 dB SPL	102 dB SPL	102 dB SPL
Full-on Gain	Peak	35 dB	35 dB	35 dB
	HF Average	30 dB	30 dB	30 dB
Reference Test Gain		26 dB	26 dB	26 dB
Frequency Range		100-9200 Hz	100-7500 Hz	100-7500 Hz
Total Harmonic Distortion	500 Hz	< 2 %	< 2 %	< 2 %
	800 Hz	< 2 %	< 2 %	< 2 %
	1600 Hz	< 2 %	< 2 %	< 2 %
Equivalent Input Noise Level	(omni/dir)	18/27 dB SPL	19/28 dB SPL	19/28 dB SPL
HF Average SPLITS	(left/right ear)	85/85 dB SPL	85/85 dB SPL	85/85 dB SPL
Attack Time		2 ms	2 ms	2 ms
Release Time		30 ms	30 ms	30 ms

0 dB SPL ref. 20 μ Pa		Oticon Opn S 1, 2, 3, Opn Play 1, 2, Ruby 1, 2
Expected operating time*	Hours	24 hrs
Latency		8.2 ms
Maximum Induction Coil Sensitivity	Measured output at 1 mA/m	56 dB SPL
	Measured output at 10 mA/m	76 dB SPL
	Measured output at 31.6 mA/m	86 dB SPL

*Expected operating time for rechargeable battery depends on use pattern, active feature set, hearing loss, sound environment, battery age and use of wireless accessories.

Technical Data

miniRITE R

85

Measured according to American National Standard ANSI S3.22-2014
and ANSI S3.55-2014/Part 5

Supply voltage: Lithium ion

0 dB SPL ref. 20 µPa		Oticon Opn S 1 Opn Play 1	Oticon Opn S 2 Ruby 1	Oticon Opn S 3 Opn Play 2 Ruby 2
OSPL90	Peak	116 dB SPL	116 dB SPL	116 dB SPL
	HF Average	112 dB SPL	112 dB SPL	112 dB SPL
Full-on Gain	Peak	54 dB	54 dB	54 dB
	HF Average	47 dB	47 dB	47 dB
Reference Test Gain		34 dB	34 dB	34 dB
Frequency Range		100-8500 Hz	100-7500 Hz	100-7500 Hz
Total Harmonic Distortion	500 Hz	< 2 %	< 2 %	< 2 %
	800 Hz	< 2 %	< 2 %	< 2 %
	1600 Hz	< 2 %	< 2 %	< 2 %
Equivalent Input Noise Level	(omni/dir)	20/29 dB SPL	21/30 dB SPL	21/30 dB SPL
HF Average SPLITS	(left/right ear)	94/94 dB SPL	94/94 dB SPL	94/94 dB SPL
Attack Time		2 ms	2 ms	2 ms
Release Time		20 ms	20 ms	20 ms

0 dB SPL ref. 20 μ Pa		Oticon Opn S 1, 2, 3, Opn Play 1, 2, Ruby 1, 2
Expected operating time*	Hours	24 hrs
Latency		8.2 ms
Maximum Induction Coil Sensitivity	Measured output at 1 mA/m	73 dB SPL
	Measured output at 10 mA/m	93 dB SPL
	Measured output at 31.6 mA/m	103 dB SPL

*Expected operating time for rechargeable battery depends on use pattern, active feature set, hearing loss, sound environment, battery age and use of wireless accessories.

Technical Data

miniRITE R **100**

Measured according to American National Standard ANSI S3.22-2014
and ANSI S3.55-2014/Part 5

Supply voltage: Lithium ion

0 dB SPL ref. 20 µPa		Oticon Opn S 1 Opn Play 1	Oticon Opn S 2 Ruby 1	Oticon Opn S 3 Opn Play 2 Ruby 2
OSPL90	Peak	122 dB SPL	122 dB SPL	122 dB SPL
	HF Average	118 dB SPL	118 dB SPL	118 dB SPL
Full-on Gain	Peak	57 dB	57 dB	57 dB
	HF Average	51 dB	51 dB	51 dB
Reference Test Gain		42 dB	42 dB	42 dB
Frequency Range		100-8000 Hz	100-7500 Hz	100-7500 Hz
Total Harmonic Distortion	500 Hz	< 2 %	< 2 %	< 2 %
	800 Hz	< 2 %	< 2 %	< 2 %
	1600 Hz	< 2 %	< 2 %	< 2 %
Equivalent Input Noise Level	(omni/dir)	19/30 dB SPL	19/30 dB SPL	19/30 dB SPL
HF Average SPLITS	(left/right ear)	103/103 dB SPL	103/103 dB SPL	103/103 dB SPL
Attack Time		2 ms	2 ms	2 ms
Release Time		10 ms	10 ms	10 ms

0 dB SPL ref. 20 μ Pa		Oticon Opn S 1, 2, 3, Opn Play 1, 2, Ruby 1, 2
Expected operating time*	Hours	24 hrs
Latency		8.2 ms
Maximum Induction Coil Sensitivity	Measured output at 1 mA/m	80 dB SPL
	Measured output at 10 mA/m	100 dB SPL
	Measured output at 31.6 mA/m	110 dB SPL

*Expected operating time for rechargeable battery depends on use pattern, active feature set, hearing loss, sound environment, battery age and use of wireless accessories.

Technical Data

Measured according to American National Standard ANSI S3.22-2014
and ANSI S3.55-2014/Part 5

Supply voltage: Lithium ion

0 dB SPL ref. 20 µPa		Oticon Opn S 1 Opn Play 1	Oticon Opn S 2 Ruby 1	Oticon Opn S 3 Opn Play 2 Ruby 2
OSPL90	Peak	127 dB SPL	127 dB SPL	127 dB SPL
	HF Average	122 dB SPL	122 dB SPL	122 dB SPL
Full-on Gain	Peak	64 dB	64 dB	64 dB
	HF Average	57 dB	57 dB	57 dB
Reference Test Gain		46 dB	46 dB	46 dB
Frequency Range		100-7800 Hz	100-6500 Hz	100-6500 Hz
Total Harmonic Distortion	500 Hz	< 2 %	< 2 %	< 2 %
	800 Hz	< 2 %	< 2 %	< 2 %
	1600 Hz	< 2 %	< 2 %	< 2 %
Equivalent Input Noise Level	(omni/dir)	18/29 dB SPL	18/29 dB SPL	18/29 dB SPL
HF Average SPLITS	(left/right ear)	105/105 dB SPL	105/105 dB SPL	105/105 dB SPL
Attack Time		2 ms	2 ms	2 ms
Release Time		20 ms	20 ms	20 ms

0 dB SPL ref. 20 μ Pa		Oticon Opn S 1, 2, 3, Opn Play 1, 2, Ruby 1, 2
Expected operating time*	Hours	24 hrs
Latency		8.2 ms
Maximum Induction Coil Sensitivity	Measured output at 1 mA/m	86 dB SPL
	Measured output at 10 mA/m	106 dB SPL
	Measured output at 31.6 mA/m	116 dB SPL

*Expected operating time for rechargeable battery depends on use pattern, active feature set, hearing loss, sound environment, battery age and use of wireless accessories.

Technical Data



miniRITE R

60

Oticon Opn S 1
Oticon Opn Play 1



Oticon Opn S 2 & Opn S 3
Oticon Opn Play 2
Ruby 1 & Ruby 2

OSPL90 - Output Sound Pressure Level
Input: 90 dB SPL
Technical setting: A0





Frequency response
Input:
60 dB SPL (acoustic)
31.6 mA/m (magnetic)
Technical setting: N0

— Acoustic input
- - - Magnetic input



Input-Output
Technical Setting: N0




Oticon Opn S 1 Oticon Opn Play 1

Oticon Opn S 2 & Opn S 3 Oticon Opn Play 2 Ruby 1 & Ruby 2

OSPL90 - Output Sound Pressure Level

Input: 90 dB SPL

Technical setting: AO



Frequency response


Input:

60 dB SPL (acoustic)

31.6 mA/m (magnetic)





Technical setting: NO

— Acoustic input
--- Magnetic input



Input-Output

Technical Setting: NO




Technical Data

miniRITE R 100

Oticon Opn S 1
Oticon Opn Play 1

Oticon Opn S 2 & Opn S 3
Oticon Opn Play 2
Ruby 1 & Ruby 2





Oticon Opn S 1 Oticon Opn Play 1

Oticon Opn S 2 & Opn S 3 Oticon Opn Play 2 Ruby 1 & Ruby 2

OSPL90 - Output Sound Pressure Level



Input: 90 dB SPL
Technical setting: A0



Frequency response



Input:
60 dB SPL (acoustic)
31.6 mA/m (magnetic)
Technical setting: N0

— Acoustic input
--- Magnetic input



Input-Output

Technical Setting: N0





0000226964000001

oticon
life-changing **technology**

22696405 / 2024.10.11 / v3